

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
 Data File : VN068073.D
 Acq On : 11 Aug 2021 15:20
 Operator : JC/MD
 Sample : M3346-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LPA-1

Manual Integrations
 APPROVED

MMDadoda
 8/12/2021 4:51:16 PM

Quant Time: Aug 12 07:12:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	291663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	580585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	618344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	218612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	235354	57.000	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 114.000%			
35) Dibromofluoromethane	8.024	113	180179	53.251	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.500%			
50) Toluene-d8	10.443	98	765685	49.109	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 98.220%			
62) 4-Bromofluorobenzene	12.736	95	255523	45.233	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.460%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	38270	25.402	ug/l	97
18) Methyl Acetate	4.870	43	5428m	1.399	ug/l	
20) Methylene Chloride	5.090	84	21261	5.861	ug/l	90
43) Isopropyl Acetate	8.568	43	38872	4.615	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
Data File : VN068073.D
Acq On : 11 Aug 2021 15:20
Operator : JC/MD
Sample : M3346-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LPA-1

Manual Integrations
APPROVED

MMDadoda
8/12/2021 4:51:16 PM

Quant Time: Aug 12 07:12:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

