

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
 Data File : VN068076.D
 Acq On : 11 Aug 2021 17:27
 Operator : JC/MD
 Sample : M3360-02 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41265

Quant Time: Aug 12 07:13:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	300702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	597217	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	583717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	203817	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	239870	56.348	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.700%
35) Dibromofluoromethane	8.021	113	184476	53.003	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.000%
50) Toluene-d8	10.443	98	759353	47.485	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.960%
62) 4-Bromofluorobenzene	12.736	95	249449	43.181	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
Data File : VN068076.D
Acq On : 11 Aug 2021 17:27
Operator : JC/MD
Sample : M3360-02 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41265

Quant Time: Aug 12 07:13:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

