

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
 Data File : VN068078.D
 Acq On : 11 Aug 2021 18:16
 Operator : JC/MD
 Sample : M3360-04 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41935

Quant Time: Aug 12 07:13:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	290231	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	566898	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	549109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	202335	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	228037	55.501	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.000%
35) Dibromofluoromethane	8.024	113	177125	53.613	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.220%
50) Toluene-d8	10.443	98	718013	47.316	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.640%
62) 4-Bromofluorobenzene	12.736	95	242157	44.048	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.100%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	11.851	91	6971	0.351	ug/l	93
68) m/p-Xylenes	11.953	106	10962	5.273	ug/l	83
69) o-Xylene	12.278	106	8821	1.206	ug/l	87
80) 1,3,5-Trimethylbenzene	13.064	105	9045	0.676	ug/l	90
84) 1,2,4-Trimethylbenzene	13.372	105	22967	1.772	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
Data File : VN068078.D
Acq On : 11 Aug 2021 18:16
Operator : JC/MD
Sample : M3360-04 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41935

Quant Time: Aug 12 07:13:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

