

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
 Data File : VN073821.D
 Acq On : 11 Aug 2022 13:35
 Operator : JC\MD
 Sample : PB146912TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146912TB

Quant Time: Aug 12 06:00:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	394409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	629984	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	635254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	306590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	225246	46.863	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.720%
35) Dibromofluoromethane	7.951	113	215823	54.557	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.120%
50) Toluene-d8	10.380	98	732127	47.031	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.060%
62) 4-Bromofluorobenzene	12.674	95	279210	51.033	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.060%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	85171	32.991	ug/l	98
20) Methylene Chloride	5.010	84	17875	2.172	ug/l	90
25) 2-Butanone	7.269	43	27553	6.872	ug/l	90
37) Ethyl Acetate	7.345	43	74068	9.975	ug/l	99
43) Isopropyl Acetate	8.498	43	283984	28.316	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
Data File : VN073821.D
Acq On : 11 Aug 2022 13:35
Operator : JC\MD
Sample : PB146912TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146912TB

Quant Time: Aug 12 06:00:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

