

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
 Data File : VN073826.D
 Acq On : 11 Aug 2022 15:38
 Operator : JC\MD
 Sample : PB146912ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146912ZHE#01

Quant Time: Aug 12 06:03:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	389196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	628159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	645758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	304398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	231213	48.748	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.500%
35) Dibromofluoromethane	7.951	113	216991	55.012	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.020%
50) Toluene-d8	10.380	98	734276	47.306	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.620%
62) 4-Bromofluorobenzene	12.674	95	284319	52.059	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.120%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	80910	31.761	ug/l	97
20) Methylene Chloride	5.022	84	17992	2.261	ug/l	98
25) 2-Butanone	7.269	43	26509	6.701	ug/l	92
37) Ethyl Acetate	7.339	43	69609	9.402	ug/l	99
43) Isopropyl Acetate	8.498	43	274653	27.465	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
Data File : VN073826.D
Acq On : 11 Aug 2022 15:38
Operator : JC\MD
Sample : PB146912ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146912ZHE#01

Quant Time: Aug 12 06:03:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

