

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
 Data File : VN073828.D
 Acq On : 11 Aug 2022 16:28
 Operator : JC\MD
 Sample : PB146912ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146912ZHE#03

Quant Time: Aug 12 06:03:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	383090	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	632028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	618121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	290364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	227535	48.738	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.480%		
35) Dibromofluoromethane	7.957	113	215412	54.277	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	108.560%		
50) Toluene-d8	10.380	98	726495	46.518	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.040%		
62) 4-Bromofluorobenzene	12.674	95	270213	49.326	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.660%		
Target Compounds					Qvalue	
16) Acetone	4.228	43	84446	33.677	ug/l	100
18) Methyl Acetate	4.793	43	7833	1.256	ug/l #	89
20) Methylene Chloride	5.016	84	17686	2.255	ug/l	98
25) 2-Butanone	7.275	43	26075	6.696	ug/l #	87
37) Ethyl Acetate	7.339	43	73089	9.812	ug/l	99
43) Isopropyl Acetate	8.498	43	277498	27.580	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
Data File : VN073828.D
Acq On : 11 Aug 2022 16:28
Operator : JC\MD
Sample : PB146912ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146912ZHE#03

Quant Time: Aug 12 06:03:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

