

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
 Data File : VN073829.D
 Acq On : 11 Aug 2022 16:53
 Operator : JC\MD
 Sample : PB146912ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146912ZHE#04

Quant Time: Aug 12 06:04:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	376942	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	623899	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	597090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	257762	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	220154	47.926	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.860%		
35) Dibromofluoromethane	7.951	113	208179	53.138	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	106.280%		
50) Toluene-d8	10.380	98	705376	45.754	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.500%		
62) 4-Bromofluorobenzene	12.674	95	250421	46.477	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.960%		
Target Compounds						
16) Acetone	4.228	43	82812	33.564	ug/l	99
18) Methyl Acetate	4.787	43	8513	1.387	ug/l	93
20) Methylene Chloride	5.016	84	18700	2.595	ug/l	91
25) 2-Butanone	7.263	43	26887	7.017	ug/l	91
37) Ethyl Acetate	7.345	43	68991	9.382	ug/l	97
43) Isopropyl Acetate	8.498	43	275549	27.743	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
Data File : VN073829.D
Acq On : 11 Aug 2022 16:53
Operator : JC\MD
Sample : PB146912ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146912ZHE#04

Quant Time: Aug 12 06:04:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

