

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
 Data File : VN073836.D
 Acq On : 11 Aug 2022 19:46
 Operator : JC\MD
 Sample : N4136-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-4

Quant Time: Aug 12 06:09:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	402386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	630456	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	599356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	291210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	229876	46.878	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.760%		
35) Dibromofluoromethane	7.951	113	82504	20.840	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	41.680%#		
50) Toluene-d8	10.381	98	704349	45.213	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.420%		
62) 4-Bromofluorobenzene	12.675	95	270956	49.570	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	99.140%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.299	59	6365	6.019	ug/l #	92
16) Acetone	4.228	43	134224	50.961	ug/l	99
18) Methyl Acetate	4.793	43	7813	1.192	ug/l #	85
20) Methylene Chloride	5.010	84	36460	6.646	ug/l	93
25) 2-Butanone	7.269	43	45926	11.228	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
Data File : VN073836.D
Acq On : 11 Aug 2022 19:46
Operator : JC\MD
Sample : N4136-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-4

Quant Time: Aug 12 06:09:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

