

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
 Data File : VN073837.D
 Acq On : 11 Aug 2022 20:11
 Operator : JC\MD
 Sample : N4136-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-5

Quant Time: Aug 12 06:10:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	395155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	629978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	627250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	309143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	227270	47.195	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.380%
35) Dibromofluoromethane	7.951	113	217503	54.982	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.960%
50) Toluene-d8	10.381	98	740662	47.580	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.160%
62) 4-Bromofluorobenzene	12.674	95	282949	51.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.360%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	198158	76.612	ug/l	99
18) Methyl Acetate	4.781	43	10667	1.658	ug/l #	87
20) Methylene Chloride	5.016	84	35708	6.622	ug/l	94
25) 2-Butanone	7.275	43	34997	8.713	ug/l	93
37) Ethyl Acetate	7.340	43	69542	9.366	ug/l	99
43) Isopropyl Acetate	8.498	43	266128	26.536	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081122\
Data File : VN073837.D
Acq On : 11 Aug 2022 20:11
Operator : JC\MD
Sample : N4136-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-5

Quant Time: Aug 12 06:10:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

