

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081123\
 Data File : VN078860.D
 Acq On : 11 Aug 2023 13:13
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Aug 12 01:40:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 09 05:05:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	28315	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	146880	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	118702	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	70418	30.891	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.967%
60) 4-Bromofluorobenzene	12.853	95	35717	21.436	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	71.467%
63) Toluene-d8	10.570	98	181057	31.127	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.767%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081123\
Data File : VN078860.D
Acq On : 11 Aug 2023 13:13
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Aug 12 01:40:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 09 05:05:47 2023
Response via : Initial Calibration

