

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081123\
 Data File : VN078869.D
 Acq On : 11 Aug 2023 17:40
 Operator : JC\MD
 Sample : 03987-02DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(JULY)DL

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/14/2023
 Supervised By :Mahesh Dadoda 08/15/2023

Quant Time: Aug 12 01:43:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 09 05:05:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.830	128	27252	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	137725	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	112556	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	68180	31.075	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	103.600%	
60) 4-Bromofluorobenzene	12.853	95	32357	20.480	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	68.267%	
63) Toluene-d8	10.577	98	164753	29.870	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	99.567%	
Target Compounds						
15) Acetone	4.448	58	11323	50.559	ug/l	86
30) 2-Butanone	7.512	43	3435m	3.132	ug/l	
62) Toluene	10.635	91	148767	22.450	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081123\
Data File : VN078869.D
Acq On : 11 Aug 2023 17:40
Operator : JC\MD
Sample : 03987-02DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JULY)DL

Quant Time: Aug 12 01:43:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 09 05:05:47 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/14/2023
Supervised By :Mahesh Dadoda 08/15/2023

