

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057204.D
 Acq On : 12 Aug 2019 9:59
 Operator : JC/SP
 Sample : VN0812WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0812WBL01

Quant Time: Aug 13 01:56:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081019W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sun Aug 11 03:53:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.58	128	181351	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.01	114	938379	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.91	117	719435	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.44	65	343546	29.06	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.87%
60) 4-Bromofluorobenzene	11.92	95	198528	18.53	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	61.77%
63) Toluene-d8	9.56	98	1091905	32.88	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	109.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057204.D
Acq On : 12 Aug 2019 9:59
Operator : JC/SP
Sample : VN0812WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0812WBL01

Quant Time: Aug 13 01:56:40 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081019W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sun Aug 11 03:53:25 2019
Response via : Initial Calibration

