

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057214.D
 Acq On : 12 Aug 2019 14:49
 Operator : JC/SP
 Sample : K4262-04DL 4X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-19D_080819DL

Quant Time: Aug 13 02:17:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1305403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1795491	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1481502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	360686	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	600865	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
35) Dibromofluoromethane	6.99	113	547815	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
50) Toluene-d8	9.56	98	2094456	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	11.92	95	541150	38.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.92%	
Target Compounds						
16) Acetone	3.49	43	7335	1.809	ug/l #	79
18) Methyl Acetate	3.90	43	19617	0.980	ug/l	97
44) Trichloroethene	8.27	130	11772	0.898	ug/l	96
64) Tetrachloroethene	10.11	164	375829	30.504	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057214.D
Acq On : 12 Aug 2019 14:49
Operator : JC/SP
Sample : K4262-04DL 4X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-19D_080819DL

Quant Time: Aug 13 02:17:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

