

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057223.D
 Acq On : 12 Aug 2019 18:27
 Operator : JC/SP
 Sample : K4229-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-C2-(0)

Quant Time: Aug 13 03:06:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1290144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1807265	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1483304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	377457	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	602067	45.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.76%	
35) Dibromofluoromethane	6.98	113	548180	48.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.68%	
50) Toluene-d8	9.56	98	2110442	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
62) 4-Bromofluorobenzene	11.92	95	539596	38.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.20%	
Target Compounds						
18) Methyl Acetate	3.90	43	26373	1.573	ug/l	93
20) Methylene Chloride	4.10	84	21960	1.661	ug/l	96
43) Isopropyl Acetate	7.57	43	153402	7.328	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057223.D
Acq On : 12 Aug 2019 18:27
Operator : JC/SP
Sample : K4229-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-C2-(0)

Quant Time: Aug 13 03:06:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

