

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057228.D
 Acq On : 12 Aug 2019 20:29
 Operator : JC/SP
 Sample : K4229-27
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-C3-(1)

Quant Time: Aug 13 03:22:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1280868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1791205	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1483895	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	365357	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	594308	45.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.24%	
35) Dibromofluoromethane	6.99	113	538625	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
50) Toluene-d8	9.56	98	2086524	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
62) 4-Bromofluorobenzene	11.92	95	541782	39.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.20%	
Target Compounds						
16) Acetone	3.48	43	50259	12.634	ug/l	95
18) Methyl Acetate	3.90	43	24399	1.420	ug/l	100
20) Methylene Chloride	4.10	84	50872	3.875	ug/l	94
43) Isopropyl Acetate	7.56	43	155977	7.518	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057228.D
Acq On : 12 Aug 2019 20:29
Operator : JC/SP
Sample : K4229-27
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-C3-(1)

Quant Time: Aug 13 03:22:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

