

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057229.D
 Acq On : 12 Aug 2019 20:53
 Operator : JC/SP
 Sample : K4229-28
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-C4-(1)

Quant Time: Aug 13 03:25:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1221599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1728788	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1469515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	380126	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	577691	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
35) Dibromofluoromethane	6.99	113	523600	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	9.56	98	2011754	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
62) 4-Bromofluorobenzene	11.92	95	535005	40.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.02%	
Target Compounds						
18) Methyl Acetate	3.90	43	24239	1.507	ug/l	99
43) Isopropyl Acetate	7.56	43	158886	7.935	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057229.D
Acq On : 12 Aug 2019 20:53
Operator : JC/SP
Sample : K4229-28
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-C4-(1)

Quant Time: Aug 13 03:25:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

