

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057230.D
 Acq On : 12 Aug 2019 21:18
 Operator : JC/SP
 Sample : K4229-29
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-A1-(0)

Quant Time: Aug 13 03:27:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1311683	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1818226	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1510346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	392213	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	604430	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
35) Dibromofluoromethane	6.99	113	551484	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	9.56	98	2134689	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	11.92	95	551782	39.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.46%	
Target Compounds						
16) Acetone	3.48	43	23174	5.688	ug/l	92
18) Methyl Acetate	3.90	43	25306	1.447	ug/l	97
20) Methylene Chloride	4.11	84	21215	1.578	ug/l #	86
43) Isopropyl Acetate	7.56	43	188907	8.970	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057230.D
Acq On : 12 Aug 2019 21:18
Operator : JC/SP
Sample : K4229-29
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-A1-(0)

Quant Time: Aug 13 03:27:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

