

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057231.D
 Acq On : 12 Aug 2019 21:42
 Operator : JC/SP
 Sample : K4229-30
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-A2-(0)

Quant Time: Aug 13 03:29:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1315772	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1821644	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1495956	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	358305	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	602655	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.08%	
35) Dibromofluoromethane	6.99	113	552124	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
50) Toluene-d8	9.56	98	2114436	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
62) 4-Bromofluorobenzene	11.92	95	530314	37.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.26%	
Target Compounds						
16) Acetone	3.48	43	48199	11.794	ug/l	91
18) Methyl Acetate	3.91	43	25536	1.459	ug/l #	71
20) Methylene Chloride	4.10	84	56252	4.171	ug/l	98
43) Isopropyl Acetate	7.56	43	191300	9.067	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057231.D
Acq On : 12 Aug 2019 21:42
Operator : JC/SP
Sample : K4229-30
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-A2-(0)

Quant Time: Aug 13 03:29:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

