

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057232.D
 Acq On : 12 Aug 2019 22:07
 Operator : JC/SP
 Sample : K4229-31
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-A3-(0)

Quant Time: Aug 13 03:31:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1292024	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1816879	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1522931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	410013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	602519	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
35) Dibromofluoromethane	6.99	113	548246	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
50) Toluene-d8	9.56	98	2097204	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
62) 4-Bromofluorobenzene	11.92	95	571035	40.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.26%	
Target Compounds						
16) Acetone	3.48	43	24018	5.985	ug/l	98
18) Methyl Acetate	3.90	43	29571	1.840	ug/l	98
43) Isopropyl Acetate	7.56	43	169444	8.052	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057232.D
Acq On : 12 Aug 2019 22:07
Operator : JC/SP
Sample : K4229-31
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-A3-(0)

Quant Time: Aug 13 03:31:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

