

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081219\
 Data File : VN057233.D
 Acq On : 12 Aug 2019 22:31
 Operator : JC/SP
 Sample : K4229-32
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-A4-(0)

Quant Time: Aug 13 03:35:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1360379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1904422	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1580223	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	389708	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	619567	44.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.56%	
35) Dibromofluoromethane	6.99	113	572599	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
50) Toluene-d8	9.56	98	2206553	48.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.52%	
62) 4-Bromofluorobenzene	11.92	95	567188	38.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.00%	
Target Compounds						
16) Acetone	3.48	43	22314	5.281	ug/l	92
18) Methyl Acetate	3.90	43	29124	1.678	ug/l	97
43) Isopropyl Acetate	7.56	43	207630	9.413	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081219\
Data File : VN057233.D
Acq On : 12 Aug 2019 22:31
Operator : JC/SP
Sample : K4229-32
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-A4-(0)

Quant Time: Aug 13 03:35:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

