

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062851.D
 Acq On : 12 Aug 2020 12:14
 Operator : JC/MD
 Sample : VN0812WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0812WBL01

Quant Time: Aug 13 05:42:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	121072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	207343	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	214897	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	79733	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	104098	51.99	ug/l	0.00
Spiked Amount						
			Recovery	=	103.98%	
35) Dibromofluoromethane	7.55	113	76452	53.90	ug/l	0.00
Spiked Amount						
			Recovery	=	107.80%	
50) Toluene-d8	10.06	98	270296	48.81	ug/l	0.00
Spiked Amount						
			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	12.38	95	98666	47.15	ug/l	0.00
Spiked Amount						
			Recovery	=	94.30%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062851.D
Acq On : 12 Aug 2020 12:14
Operator : JC/MD
Sample : VN0812WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0812WBL01

Quant Time: Aug 13 05:42:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

