

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062856.D
 Acq On : 12 Aug 2020 14:31
 Operator : JC/MD
 Sample : L3666-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB179-GW-598-600

Quant Time: Aug 13 05:57:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	122400	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	214343	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	215246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80952	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	106126	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
35) Dibromofluoromethane	7.55	113	74042	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
50) Toluene-d8	10.06	98	272771	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
62) 4-Bromofluorobenzene	12.38	95	104136	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
Target Compounds						
16) Acetone	3.79	43	6463	8.837	ug/l	93
52) Toluene	10.13	92	2324	0.604	ug/l	95
95) Naphthalene	15.11	128	3320	4.534	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062856.D
Acq On : 12 Aug 2020 14:31
Operator : JC/MD
Sample : L3666-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB179-GW-598-600

Quant Time: Aug 13 05:57:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

