

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062865.D
 Acq On : 12 Aug 2020 18:36
 Operator : JC/MD
 Sample : PB130874TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130874TB

Quant Time: Aug 13 06:24:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	125840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	218568	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	233101	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105269	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
35) Dibromofluoromethane	7.55	113	76439	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
50) Toluene-d8	10.06	98	295062	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
62) 4-Bromofluorobenzene	12.38	95	110774	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
Target Compounds						
16) Acetone	3.78	43	16461	21.891	ug/l	98
20) Methylene Chloride	4.50	84	8151	5.243	ug/l	91
43) Isopropyl Acetate	8.13	43	40704	10.977	ug/l	98
95) Naphthalene	15.11	128	8070	5.545	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062865.D
Acq On : 12 Aug 2020 18:36
Operator : JC/MD
Sample : PB130874TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130874TB

Quant Time: Aug 13 06:24:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

