

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062878.D
 Acq On : 13 Aug 2020 00:42
 Operator : JC/MD
 Sample : PB130908TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130908TB

Quant Time: Aug 13 07:13:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	137316	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	236969	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	237904	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88333	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	112125	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	7.55	113	82400	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
50) Toluene-d8	10.06	98	304940	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
62) 4-Bromofluorobenzene	12.38	95	110097	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
Target Compounds						
16) Acetone	3.78	43	4290	5.228	ug/l #	88
20) Methylene Chloride	4.50	84	7079	4.015	ug/l	91
95) Naphthalene	15.11	128	4182	4.662	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062878.D
Acq On : 13 Aug 2020 00:42
Operator : JC/MD
Sample : PB130908TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130908TB

Quant Time: Aug 13 07:13:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

