

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062888.D
 Acq On : 13 Aug 2020 4:45
 Operator : JC/MD
 Sample : PB130908ZHE#18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130908ZHE#18

Quant Time: Aug 13 07:45:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	125275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	213412	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	219165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	79790	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	103093	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
35) Dibromofluoromethane	7.55	113	74744	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	10.06	98	278152	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	12.38	95	100783	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
Target Compounds						
16) Acetone	3.78	43	3744	5.002	ug/l	96
20) Methylene Chloride	4.50	84	6755	4.235	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062888.D
Acq On : 13 Aug 2020 4:45
Operator : JC/MD
Sample : PB130908ZHE#18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130908ZHE#18

Quant Time: Aug 13 07:45:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

