

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062890.D
 Acq On : 13 Aug 2020 5:33
 Operator : JC/MD
 Sample : L3648-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BS-STOCK-1

Quant Time: Aug 13 07:51:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	122766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	211724	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	217287	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80595	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	107292	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
35) Dibromofluoromethane	7.55	113	74853	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
50) Toluene-d8	10.06	98	280469	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
62) 4-Bromofluorobenzene	12.38	95	100672	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
Target Compounds						
16) Acetone	3.78	43	5661	7.717	ug/l #	85
20) Methylene Chloride	4.50	84	9828	6.662	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062890.D
Acq On : 13 Aug 2020 5:33
Operator : JC/MD
Sample : L3648-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BS-STOCK-1

Quant Time: Aug 13 07:51:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

