

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062891.D
 Acq On : 13 Aug 2020 5:57
 Operator : JC/MD
 Sample : L3648-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BS-STOCK-2

Quant Time: Aug 13 07:53:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	131159	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	229819	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	209534	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	78355	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	112165	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
35) Dibromofluoromethane	7.55	113	83392	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
50) Toluene-d8	10.06	98	272473	44.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.78%	
62) 4-Bromofluorobenzene	12.38	95	97545	42.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.12%	
Target Compounds						
16) Acetone	3.78	43	6211	7.925	ug/l	95
20) Methylene Chloride	4.50	84	7655	4.648	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062891.D
Acq On : 13 Aug 2020 5:57
Operator : JC/MD
Sample : L3648-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BS-STOCK-2

Quant Time: Aug 13 07:53:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

