

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062892.D
 Acq On : 13 Aug 2020 6:22
 Operator : JC/MD
 Sample : L3648-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BS-STOCK-3

Quant Time: Aug 13 07:54:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	121967	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	211250	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	213321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80081	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	104722	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
35) Dibromofluoromethane	7.55	113	73418	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
50) Toluene-d8	10.06	98	273333	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
62) 4-Bromofluorobenzene	12.38	95	97823	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
Target Compounds						
16) Acetone	3.79	43	4528	6.213	ug/l	99
20) Methylene Chloride	4.50	84	6715	4.340	ug/l	89
30) Chloroform	7.33	83	8025	2.835	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062892.D
Acq On : 13 Aug 2020 6:22
Operator : JC/MD
Sample : L3648-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BS-STOCK-3

Quant Time: Aug 13 07:54:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

