

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062895.D
 Acq On : 13 Aug 2020 7:34
 Operator : JC/MD
 Sample : L3648-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BS-STOCK-6

Quant Time: Aug 13 07:59:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	151102	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	267874	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	274827	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	104141	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	127544	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
35) Dibromofluoromethane	7.55	113	93816	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	10.06	98	345332	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
62) 4-Bromofluorobenzene	12.38	95	128421	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
Target Compounds						
16) Acetone	3.78	43	6277	6.952	ug/l	98
20) Methylene Chloride	4.50	84	7895	4.080	ug/l	89
30) Chloroform	7.32	83	8889	2.534	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062895.D
Acq On : 13 Aug 2020 7:34
Operator : JC/MD
Sample : L3648-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BS-STOCK-6

Quant Time: Aug 13 07:59:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

