

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062896.D
 Acq On : 13 Aug 2020 7:59
 Operator : JC/MD
 Sample : L3648-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BS-STOCK-7

Quant Time: Aug 13 08:24:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	130907	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	234322	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	240125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	95569	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	113629	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
35) Dibromofluoromethane	7.55	113	82213	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
50) Toluene-d8	10.06	98	304990	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
62) 4-Bromofluorobenzene	12.38	95	116258	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
Target Compounds						
16) Acetone	3.79	43	5575	7.127	ug/l	98
20) Methylene Chloride	4.50	84	7477	4.532	ug/l	94
30) Chloroform	7.33	83	7946	2.615	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062896.D
Acq On : 13 Aug 2020 7:59
Operator : JC/MD
Sample : L3648-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BS-STOCK-7

Quant Time: Aug 13 08:24:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

