

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081220\
 Data File : VN062851.D
 Acq On : 12 Aug 2020 12:14
 Operator : JC/MD
 Sample : VN0812WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0812WBL01

Quant Time: Aug 13 05:42:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	121072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	207343	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	214897	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	79733	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	104098	51.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.98%	
35) Dibromofluoromethane	7.55	113	76452	53.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.80%	
50) Toluene-d8	10.06	98	270296	48.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.62%	
62) 4-Bromofluorobenzene	12.38	95	98666	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081220\
Data File : VN062851.D
Acq On : 12 Aug 2020 12:14
Operator : JC/MD
Sample : VN0812WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0812WBL01

Quant Time: Aug 13 05:42:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.00 min

Lab File: VN062851.D

Acq: 12 Aug 2020 12:14

Instrument :

MSVOA_N

ClientSampleId :


VN0812WBL01

Tot Ion:168 Resp: 121072

Ion Ratio Lower Upper

168 100

99 63.7 52.9 79.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.63

60000

50000

40000



30000

20000

10000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 51.993 ug/l

RT: 7.99 min Scan# 1933

Delta R.T. -0.00 min

Lab File: VN062851.D


Acq: 12 Aug 2020 12:14

Tgt Ion: 65 Resp: 104098

Ion Ratio Lower Upper

65 100

67 49.4 0.0 101.4



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0

7.99

50000

40000


30000


20000

10000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2108
 Delta R.T. -0.00 min
 Lab File: VN062851.D
 Acq: 12 Aug 2020 12:14

Instrument :
 MSVOA_N
 ClientSampled :
 VN0812WBL01

Tot Ion:114 Resp: 207343
 Ion Ratio Lower Upper
 114 100
 63 22.5 0.0 47.6
 88 17.7 0.0 34.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 53.902 ug/l
 RT: 7.55 min Scan# 1796
 Delta R.T. -0.00 min
 Lab File: VN062851.D
 Acq: 12 Aug 2020 12:14

Tgt Ion:113 Resp: 76452
 Ion Ratio Lower Upper
 113 100
 111 100.0 81.8 122.6
 192 18.1 15.0 22.4



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.813 ug/l

RT: 10.06 min Scan# 2578

Delta R.T. -0.00 min

Lab File: VN062851.D

Acq: 12 Aug 2020 12:14

Instrument :

MSVOA_N

ClientSampleId :





VN0812WBL01

Tot Ion: 98 Resp: 270296

Ion Ratio Lower Upper

98 100

100 63.1 50.9 76.3



#62

4-Bromofluorobenzene

Concen: 47.151 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.00 min

Lab File: VN062851.D

Acq: 12 Aug 2020 12:14




Tgt Ion: 95 Resp: 98666


Ion Ratio Lower Upper

95 100

174 74.7 0.0 144.8

176 72.3 0.0 142.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. -0.00 min
Lab File: VN062851.D
Acq: 12 Aug 2020 12:14

Instrument :
MSVOA_N
ClientSampleId :
VN0812WBL01

Tot Ion:117 Resp: 214897
Ion Ratio Lower Upper
117 100
82 59.7 49.0 73.4
119 32.0 26.2 39.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3591
Delta R.T. -0.00 min
Lab File: VN062851.D
Acq: 12 Aug 2020 12:14

Tgt Ion:152 Resp: 79733
Ion Ratio Lower Upper
152 100
115 63.4 31.6 94.8
150 155.3 0.0 349.2

