

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
 Data File : VN068084.D
 Acq On : 12 Aug 2021 12:09
 Operator : JC/MD
 Sample : VN0812WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0812WBL01

Quant Time: Aug 13 05:40:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	316077	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	602366	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	589244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	194307	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	241686	54.013	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.020%	
35) Dibromofluoromethane	8.024	113	182777	52.066	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.140%	
50) Toluene-d8	10.443	98	767332	47.566	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.140%	
62) 4-Bromofluorobenzene	12.734	95	250411	43.001	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 86.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
Data File : VN068084.D
Acq On : 12 Aug 2021 12:09
Operator : JC/MD
Sample : VN0812WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0812WBL01

Quant Time: Aug 13 05:40:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

