

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
 Data File : VN068093.D
 Acq On : 12 Aug 2021 16:40
 Operator : JC/MD
 Sample : M3361-07 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41907

Quant Time: Aug 13 05:44:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	307652	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	583163	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	571002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	200436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229125	52.608	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.220%
35) Dibromofluoromethane	8.024	113	179489	52.813	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.620%
50) Toluene-d8	10.443	98	741188	47.467	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.940%
62) 4-Bromofluorobenzene	12.733	95	252414	44.567	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.140%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	25439	16.008	ug/l	# 84
84) 1,2,4-Trimethylbenzene	13.369	105	4941	0.385	ug/l	90
95) Naphthalene	15.509	128	6573	3.857	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
Data File : VN068093.D
Acq On : 12 Aug 2021 16:40
Operator : JC/MD
Sample : M3361-07 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41907

Quant Time: Aug 13 05:44:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

