

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
 Data File : VN068094.D
 Acq On : 12 Aug 2021 17:04
 Operator : JC/MD
 Sample : M3361-08 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41908

Manual Integrations
 APPROVED

MMDadoda
 8/13/2021 10:56:35 AM

Quant Time: Aug 13 05:44:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	317499	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	609363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	594183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	211801	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	238448	53.050	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 106.100%		
35) Dibromofluoromethane	8.024	113	187858	52.899	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 105.800%		
50) Toluene-d8	10.443	98	768260	47.117	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 94.240%		
62) 4-Bromofluorobenzene	12.733	95	256363	43.458	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 86.920%		
Target Compounds						Qvalue
16) Acetone	4.301	43	4645m	2.832	ug/l	
52) Toluene	10.507	92	3227	0.298	ug/l	86
95) Naphthalene	15.507	128	9756	4.156	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
Data File : VN068094.D
Acq On : 12 Aug 2021 17:04
Operator : JC/MD
Sample : M3361-08 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41908

Manual Integrations
APPROVED

MMDadoda
8/13/2021 10:56:35 AM

Quant Time: Aug 13 05:44:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

