

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
 Data File : VN068095.D
 Acq On : 12 Aug 2021 17:29
 Operator : JC/MD
 Sample : M3361-09 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41911

Quant Time: Aug 13 05:45:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	309660	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	603381	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	593584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	203901	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	238002	54.292	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.580%
35) Dibromofluoromethane	8.024	113	183831	52.278	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.560%
50) Toluene-d8	10.443	98	767042	47.476	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.960%
62) 4-Bromofluorobenzene	12.733	95	253557	43.414	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081221\
Data File : VN068095.D
Acq On : 12 Aug 2021 17:29
Operator : JC/MD
Sample : M3361-09 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41911

Quant Time: Aug 13 05:45:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

