

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
 Data File : VN083231.D
 Acq On : 12 Aug 2024 12:35
 Operator : JC\MD
 Sample : P3506-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT4534

Quant Time: Aug 13 04:16:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306010	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	323138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146363	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	127595	55.723	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.440%
35) Dibromofluoromethane	8.171	113	101185	52.975	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.960%
50) Toluene-d8	10.565	98	391269	54.915	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.840%
62) 4-Bromofluorobenzene	12.847	95	166836	60.062	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	120.120%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	25560	28.527	ug/l #	88
20) Methylene Chloride	5.277	84	13347	6.473	ug/l #	83
25) 2-Butanone	7.483	43	17651	12.830	ug/l	91
43) Isopropyl Acetate	8.688	43	131342	22.080	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
Data File : VN083231.D
Acq On : 12 Aug 2024 12:35
Operator : JC\MD
Sample : P3506-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT4534

Quant Time: Aug 13 04:16:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

