

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
 Data File : VN083235.D
 Acq On : 12 Aug 2024 14:11
 Operator : JC\MD
 Sample : PB162664ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162664ZHE#02

Quant Time: Aug 13 04:18:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	130117	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	256396	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123688	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111782	60.355	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.720%	
35) Dibromofluoromethane	8.171	113	89607	55.991	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.980%	
50) Toluene-d8	10.565	98	345411	57.860	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 115.720%#	
62) 4-Bromofluorobenzene	12.847	95	146042	62.750	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 125.500%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	36545	50.427	ug/l	95
20) Methylene Chloride	5.271	84	11139	6.679	ug/l #	76
25) 2-Butanone	7.494	43	11538	10.369	ug/l	97
43) Isopropyl Acetate	8.688	43	166224	34.483	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
Data File : VN083235.D
Acq On : 12 Aug 2024 14:11
Operator : JC\MD
Sample : PB162664ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162664ZHE#02

Quant Time: Aug 13 04:18:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

