

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
 Data File : VN083236.D
 Acq On : 12 Aug 2024 14:35
 Operator : JC\MD
 Sample : PB162664ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162664ZHE#03

Quant Time: Aug 13 04:19:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	264296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	277683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115677	58.798	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.600%			
35) Dibromofluoromethane	8.165	113	91046	55.190	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.380%			
50) Toluene-d8	10.565	98	351359	57.097	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 114.200%#			
62) 4-Bromofluorobenzene	12.847	95	148110	61.737	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 123.480%			
Target Compounds						
16) Acetone	4.442	43	38395	49.875	ug/l	99
20) Methylene Chloride	5.277	84	11165	6.302	ug/l	85
25) 2-Butanone	7.488	43	12133	10.265	ug/l	97
43) Isopropyl Acetate	8.688	43	167925	33.744	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
Data File : VN083236.D
Acq On : 12 Aug 2024 14:35
Operator : JC\MD
Sample : PB162664ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162664ZHE#03

Quant Time: Aug 13 04:19:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

