

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
 Data File : VN083237.D
 Acq On : 12 Aug 2024 14:59
 Operator : JC\MD
 Sample : PB162664ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162664ZHE#04

Quant Time: Aug 13 04:19:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	140588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	269660	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	119150	59.542	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.080%
35) Dibromofluoromethane	8.171	113	94349	56.055	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.100%
50) Toluene-d8	10.571	98	358771	57.142	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	114.280%#
62) 4-Bromofluorobenzene	12.847	95	150532	61.498	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	123.000%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	38043	48.584	ug/l	98
20) Methylene Chloride	5.277	84	11327	6.285	ug/l	89
25) 2-Butanone	7.494	43	12411	10.323	ug/l	99
43) Isopropyl Acetate	8.688	43	175749	34.679	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
Data File : VN083237.D
Acq On : 12 Aug 2024 14:59
Operator : JC\MD
Sample : PB162664ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162664ZHE#04

Quant Time: Aug 13 04:19:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

