

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
 Data File : VN083241.D
 Acq On : 12 Aug 2024 16:36
 Operator : JC\MD
 Sample : P3530-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-A-05-202408

Quant Time: Aug 13 04:22:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	136029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262728	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117085	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	110618	57.131	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.260%
35) Dibromofluoromethane	8.171	113	85361	52.053	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.100%
50) Toluene-d8	10.565	98	322440	52.710	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.420%
62) 4-Bromofluorobenzene	12.847	95	131198	55.014	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	110.020%
Target Compounds						
30) Chloroform	7.971	83	66769	22.013	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081224\
Data File : VN083241.D
Acq On : 12 Aug 2024 16:36
Operator : JC\MD
Sample : P3530-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-A-05-202408

Quant Time: Aug 13 04:22:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

