

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081318\
 Data File : VN050569.D
 Acq On : 13 Aug 2018 15:33
 Operator : MD\SY
 Sample : PB111961ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111961ZHE#02

Quant Time: Aug 14 01:50:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	657076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1075502	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	923754	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	341019	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	447860	51.01	ug/l	0.00
Spiked Amount						
			Recovery	=	102.02%	
35) Dibromofluoromethane	7.59	113	413385	47.90	ug/l	0.00
Spiked Amount						
			Recovery	=	95.80%	
50) Toluene-d8	10.09	98	1481768	45.67	ug/l	0.00
Spiked Amount						
			Recovery	=	91.34%	
62) 4-Bromofluorobenzene	12.40	95	408246	36.92	ug/l	0.00
Spiked Amount						
			Recovery	=	73.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	14292	5.11	ug/l	89
18) Methyl Acetate	4.33	43	10675	1.38	ug/l	# 78
43) Isopropyl Acetate	8.17	43	220367	15.00	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081318\
Data File : VN050569.D
Acq On : 13 Aug 2018 15:33
Operator : MD\SY
Sample : PB111961ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111961ZHE#02

Quant Time: Aug 14 01:50:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

