

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081318\
 Data File : VN050574.D
 Acq On : 13 Aug 2018 17:36
 Operator : MD\SY
 Sample : PB111961ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111961ZHE#07

Quant Time: Aug 14 02:06:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	626778	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1046398	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	902761	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	330461	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	431106	51.48	ug/l	0.00
Spiked Amount			Recovery	=	102.96%	
35) Dibromofluoromethane	7.59	113	403864	48.10	ug/l	0.00
Spiked Amount			Recovery	=	96.20%	
50) Toluene-d8	10.09	98	1434043	45.43	ug/l	0.00
Spiked Amount			Recovery	=	90.86%	
62) 4-Bromofluorobenzene	12.40	95	401582	37.32	ug/l	0.00
Spiked Amount			Recovery	=	74.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	13733	5.15	ug/l	97
18) Methyl Acetate	4.34	43	14742	2.09	ug/l	96
43) Isopropyl Acetate	8.17	43	231477	16.19	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081318\
Data File : VN050574.D
Acq On : 13 Aug 2018 17:36
Operator : MD\SY
Sample : PB111961ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111961ZHE#07

Quant Time: Aug 14 02:06:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

