

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081318\
 Data File : VN050576.D
 Acq On : 13 Aug 2018 18:26
 Operator : MD\SY
 Sample : PB111961ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111961ZHE#09

Quant Time: Aug 14 02:11:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	628374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1055585	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	916306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	344377	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	438389	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
35) Dibromofluoromethane	7.59	113	406712	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
50) Toluene-d8	10.09	98	1435569	45.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.16%	
62) 4-Bromofluorobenzene	12.40	95	408259	37.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.22%	
Target Compounds						
16) Acetone	3.83	43	12459	4.62	ug/l	99
18) Methyl Acetate	4.34	43	14439	2.03	ug/l #	87
43) Isopropyl Acetate	8.17	43	242096	16.79	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081318\
Data File : VN050576.D
Acq On : 13 Aug 2018 18:26
Operator : MD\SY
Sample : PB111961ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111961ZHE#09

Quant Time: Aug 14 02:11:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

