

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081319\
 Data File : VN057249.D
 Acq On : 13 Aug 2019 16:10
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Aug 14 06:29:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 06:25:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1204848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1739067	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1553742	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	667608	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	610400	50.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.44%	
35) Dibromofluoromethane	7.58	113	556875	50.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.30%	
50) Toluene-d8	10.09	98	2110835	52.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.24%	
62) 4-Bromofluorobenzene	12.40	95	718649	51.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.48%	

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.85	85	498521	49.083	ug/l	100
3) Chloromethane	2.05	50	546646	47.474	ug/l	97
4) Vinyl Chloride	2.18	62	558962	49.975	ug/l	98
5) Bromomethane	2.55	94	425291	53.952	ug/l	99
6) Chloroethane	2.69	64	373094	51.771	ug/l	99
7) Trichlorofluoromethane	3.02	101	878650	49.929	ug/l	98
8) Diethyl Ether	3.40	74	330753	51.868	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	3.75	101	531112	50.262	ug/l	99
10) Methyl Iodide	3.95	142	841903	51.287	ug/l	99
11) Tert butyl alcohol	4.76	59	421152	257.305	ug/l #	82
12) 1,1-Dichloroethene	3.73	96	526835	50.249	ug/l	98
13) Acrolein	3.60	56	283311	241.934	ug/l	99
14) Allyl chloride	4.31	41	831223	52.263	ug/l	99
15) Acrylonitrile	4.97	53	1206589	261.661	ug/l	99
16) Acetone	3.81	43	1091696	261.988	ug/l	99
17) Carbon Disulfide	4.05	76	1438177	50.184	ug/l	97
18) Methyl Acetate	4.31	43	556326	51.832	ug/l	100
19) Methyl tert-butyl Ether	5.03	73	1623637	51.862	ug/l	100
20) Methylene Chloride	4.54	84	588905	48.381	ug/l	96
21) trans-1,2-Dichloroethene	5.03	96	570864	51.077	ug/l	96
22) Diisopropyl ether	5.94	45	1617139	51.713	ug/l	99
23) Vinyl Acetate	5.88	43	6539253	263.891	ug/l	99
24) 1,1-Dichloroethane	5.84	63	1005137	51.018	ug/l	100
25) 2-Butanone	6.82	43	1570984	265.172	ug/l	99
26) 2,2-Dichloropropane	6.82	77	846587	52.375	ug/l	99
27) cis-1,2-Dichloroethene	6.82	96	659875	51.200	ug/l	99
28) Bromochloromethane	7.18	49	432937	49.798	ug/l	99
29) Tetrahydrofuran	7.20	42	977686	258.150	ug/l	100
30) Chloroform	7.36	83	1021052	49.613	ug/l	98
31) Cyclohexane	7.65	56	862445	50.087	ug/l	99
32) 1,1,1-Trichloroethane	7.56	97	901654	51.795	ug/l	99
36) 1,1-Dichloropropene	7.79	75	792650	52.133	ug/l	99
37) Ethyl Acetate	6.91	43	624196	51.077	ug/l	98
38) Carbon Tetrachloride	7.77	117	810557	50.185	ug/l	100

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081319\
 Data File : VN057249.D
 Acq On : 13 Aug 2019 16:10
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Aug 14 06:29:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 06:25:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.08	83	935675	51.977	ug/l	98
40) Benzene	8.03	78	2342845	50.721	ug/l	98
41) Methacrylonitrile	7.16	41	266859	53.062	ug/l #	100
42) 1,2-Dichloroethane	8.12	62	744750	50.315	ug/l	99
43) Isopropyl Acetate	8.15	43	1060991	52.717	ug/l	100
44) Trichloroethene	8.83	130	684429	51.117	ug/l	97
45) 1,2-Dichloropropane	9.11	63	609636	52.097	ug/l	99
46) Dibromomethane	9.20	93	392857	49.792	ug/l	98
47) Bromodichloromethane	9.40	83	796352	52.352	ug/l	100
48) Methyl methacrylate	9.19	41	503660	54.084	ug/l	99
49) 1,4-Dioxane	9.19	88	204469	974.282	ug/l	97
51) 4-Methyl-2-Pentanone	9.98	43	3023004	255.792	ug/l	99
52) Toluene	10.15	92	1484164	51.720	ug/l	99
53) t-1,3-Dichloropropene	10.38	75	866942	53.282	ug/l	100
54) cis-1,3-Dichloropropene	9.83	75	982041	53.553	ug/l	99
55) 1,1,2-Trichloroethane	10.56	97	567646	50.608	ug/l	99
56) Ethyl methacrylate	10.43	69	830335	52.764	ug/l	99
57) 1,3-Dichloropropane	10.70	76	945576	51.096	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.69	63	1183610	256.695	ug/l	100
59) 2-Hexanone	10.75	43	2197622	262.082	ug/l	100
60) Dibromochloromethane	10.90	129	655523	52.115	ug/l	99
61) 1,2-Dibromoethane	11.00	107	603009	51.784	ug/l	100
64) Tetrachloroethene	10.63	164	674392	49.140	ug/l	99
65) Chlorobenzene	11.43	112	1624698	51.101	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.51	131	623468	50.780	ug/l	99
67) Ethyl Benzene	11.51	91	2845953	52.433	ug/l	99
68) m/p-Xylenes	11.62	106	2159443	106.632	ug/l	100
69) o-Xylene	11.95	106	1036298	52.263	ug/l	100
70) Styrene	11.96	104	1738870	56.642	ug/l	99
71) Bromoform	12.13	173	501492	53.305	ug/l #	100
73) Isopropylbenzene	12.25	105	2817171	54.929	ug/l	99
74) N-amyl acetate	12.07	43	840863	53.661	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.50	83	759636	54.541	ug/l	99
76) 1,2,3-Trichloropropane	12.55	75	958166	84.285	ug/l #	100
77) Bromobenzene	12.53	156	743767	53.679	ug/l	100
78) n-propylbenzene	12.59	91	3039759	54.283	ug/l	100
79) 2-Chlorotoluene	12.67	91	1836809	54.508	ug/l	100
80) 1,3,5-Trimethylbenzene	12.73	105	2375553	54.567	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.30	75	242637	48.817	ug/l	98
82) 4-Chlorotoluene	12.77	91	1786577	49.068	ug/l	99
83) tert-Butylbenzene	12.99	119	2108015	54.786	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	2264780	48.174	ug/l	99
85) sec-Butylbenzene	13.17	105	2655698	54.091	ug/l	100
86) p-Isopropyltoluene	13.29	119	2358001	48.032	ug/l	100
87) 1,3-Dichlorobenzene	13.28	146	1186971	51.322	ug/l	100
88) 1,4-Dichlorobenzene	13.36	146	1120929	49.830	ug/l	99
89) n-Butylbenzene	13.62	91	1738175	50.478	ug/l	99
90) Hexachloroethane	13.88	117	427648	52.350	ug/l	100
91) 1,2-Dichlorobenzene	13.65	146	1169889	51.881	ug/l	100
92) 1,2-Dibromo-3-Chloropropan	14.27	75	120138	53.872	ug/l	100


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081319\
Data File : VN057249.D
Acq On : 13 Aug 2019 16:10
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Quant Time: Aug 14 06:29:31 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 06:25:03 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.91	180	518447	58.412	ug/l	100
94) Hexachlorobutadiene	15.01	225	489340	57.525	ug/l	98
95) Naphthalene	15.13	128	892186	60.483	ug/l	99
96) 1,2,3-Trichlorobenzene	15.30	180	493037	61.401	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1830

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tgt Ion:168 Resp: 1204848

Ion Ratio Lower Upper




168 100

99 47.7 38.6 58.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#2

Dichlorodifluoromethane

Concen: 49.083 ug/l

RT: 1.85 min Scan# 24

Delta R.T. 0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 85 Resp: 498521

Ion Ratio Lower Upper



85 100


87 32.1 16.2 48.6



Abundance Ion 84.90 (84.60 to 85.60): VNO

Ion 86.90 (86.60 to 87.60): VNO





#3

Chloromethane

Concen: 47.474 ug/l

RT: 2.05 min Scan# 86

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 50 Resp: 546646

Ion Ratio Lower Upper

50 100

52 33.6 25.6 38.4



#4

Vinyl Chloride

Concen: 49.975 ug/l

RT: 2.18 min Scan# 127

Delta R.T. 0.00 min

Lab File: VN057249.D




Acq: 13 Aug 2019 16:10


Tgt Ion: 62 Resp: 558962

Ion Ratio Lower Upper

62 100

64 31.8 26.2 39.2





#5

Bromomethane

Concen: 53.952 ug/l

RT: 2.55 min Scan# 240

Delta R.T. 0.01 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tgt Ion: 94 Resp: 425291

Ion Ratio Lower Upper

94 100

96 95.4 75.7 113.5



Abundance Ion 93.90 (93.60 to 94.60): VN057249.D (-145) (-)

Ion 95.90 (95.60 to 96.60): VN057249.D (-145) (-)

2.55

200000

150000



100000

50000

0

2.50 2.60 2.70

Time-->



#6

Chloroethane

Concen: 51.771 ug/l

RT: 2.69 min Scan# 286

Delta R.T. 0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 64 Resp: 373094

Ion Ratio Lower Upper

64 100

66 32.3 26.3 39.5



Abundance Ion 63.90 (63.60 to 64.60): VN057249.D (-193) (-)

Ion 65.90 (65.60 to 66.60): VN057249.D (-193) (-)

2.69

200000

150000


100000


50000

0

2.60 2.70 2.80 2.90

Time-->









#7

Trichlorofluoromethane
 Concen: 49.929 ug/l
 RT: 3.02 min Scan# 387
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 101 Resp: 878650
 Ion Ratio Lower Upper
 101 100
 103 66.3 51.8 77.6




#8

Diethyl Ether
 Concen: 51.868 ug/l
 RT: 3.40 min Scan# 507
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 74 Resp: 330753
 Ion Ratio Lower Upper
 74 100
 45 89.0 45.1 135.2





#9

1,1,2-Trichlorotrifluoroethane

Concen: 50.262 ug/l

RT: 3.75 min Scan# 616

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion:101 Resp: 531112

Ion Ratio Lower Upper

101 100

85 41.1 32.7 49.1




151 88.4 70.1 105.1



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): V

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 51.287 ug/l

RT: 3.95 min Scan# 676

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10


Tgt Ion:142 Resp: 841903

Ion Ratio Lower Upper

142 100

127 39.0 30.7 46.1



141 14.0 11.0 16.6




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 257.305 ug/l
 RT: 4.76 min Scan# 930
 Delta R.T. 0.01 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Instrument :
 MSVOA_N
 ClientSampled :
 VSTDCCC050




Tgt Ion: 59 Resp: 421152
 Ion Ratio Lower Upper
 59 100
 57 10.1 3.4 5.2#




#12

1,1-Dichloroethene
 Concen: 50.249 ug/l
 RT: 3.73 min Scan# 610
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 96 Resp: 526835
 Ion Ratio Lower Upper
 96 100
 61 154.5 121.1 181.7
 98 64.9 52.0 78.0





#13

Acrolein

Concen: 241.934 ug/l

RT: 3.60 min Scan# 568

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 56 Resp: 283311

Ion Ratio Lower Upper

56 100

55 71.4 56.6 84.8



#14

Allyl chloride

Concen: 52.263 ug/l

RT: 4.31 min Scan# 790

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10




Tgt Ion: 41 Resp: 831223


Ion Ratio Lower Upper

41 100

39 60.0 48.5 72.7

76 37.2 29.8 44.8





#15

Acrylonitrile

Concen: 261.661 ug/l

RT: 4.97 min Scan# 995

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion: 53 Resp: 1206589

Ion Ratio Lower Upper

53 100

52 82.1 65.9 98.9




51 35.3 27.8 41.6



Abundance Ion 53.00 (52.70 to 53.70): VN0

Ion 52.00 (51.70 to 52.70): VN0

Ion 51.00 (50.70 to 51.70): VN0



#16

Acetone

Concen: 261.988 ug/l

RT: 3.81 min Scan# 632

Delta R.T. 0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 43 Resp: 1091696

Ion Ratio Lower Upper



43 100


58 32.6 26.7 40.1



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0





#17

Carbon Disulfide

Concen: 50.184 ug/l

RT: 4.05 min Scan# 708

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tgt Ion: 76 Resp: 1438177

Ion Ratio Lower Upper

76 100

78 8.7 7.7 11.5



Abundance Ion 75.90 (75.60 to 76.60): VN057249.D (-615) (-)

Ion 77.90 (77.60 to 78.60): VN057249.D (-615) (-)

4.05

600000

500000

400000



300000

200000

100000

0

Time-->



#18

Methyl Acetate

Concen: 51.832 ug/l

RT: 4.31 min Scan# 789

Delta R.T. 0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 43 Resp: 556326

Ion Ratio Lower Upper

43 100

74 25.7 20.6 31.0



Abundance Ion 43.00 (42.70 to 43.70): VN057249.D (-696) (-)

Ion 74.00 (73.70 to 74.70): VN057249.D (-696) (-)

4.31

200000


150000


100000

50000

0

Time-->









#19
Methyl tert-butyl Ether
Concen: 51.862 ug/l
RT: 5.03 min Scan# 1012
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 1623637
Ion Ratio Lower Upper
73 100
57 21.6 17.1 25.7



#20
Methylene Chloride
Concen: 48.381 ug/l
RT: 4.54 min Scan# 861
Delta R.T. -0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion: 84 Resp: 588905
Ion Ratio Lower Upper
84 100
49 113.8 85.9 128.9
51 35.2 27.1 40.7
86 65.0 51.0 76.6





#21

trans-1,2-Dichloroethene

Concen: 51.077 ug/l

RT: 5.03 min Scan# 1013

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion: 96 Resp: 570864

Ion Ratio Lower Upper

96 100

61 125.9 104.6 156.8

98 64.2 49.7 74.5






Abundance

Ion 95.90 (95.60 to 96.60): VN0

Ion 60.90 (60.60 to 61.60): VN0

Ion 97.90 (97.60 to 98.60): VN0



#22

Diisopropyl ether

Concen: 51.713 ug/l

RT: 5.94 min Scan# 1295

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Tgt Ion: 45 Resp: 1617139


Ion Ratio Lower Upper

45 100

43 61.5 48.9 73.3

87 31.7 25.5 38.3

59 12.6 9.7 14.5





Abundance


Ion 45.00 (44.70 to 45.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 87.00 (86.70 to 87.70): VN0

Ion 59.00 (58.70 to 59.70): VN0





#23

Vinyl Acetate

Concen: 263.891 ug/l

RT: 5.88 min Scan# 1277

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tgt Ion: 43 Resp: 6539253

Ion Ratio Lower Upper




43 100

86 11.3 9.2 13.8



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 86.00 (85.70 to 86.70): VN0



#24

1,1-Dichloroethane

Concen: 51.018 ug/l

RT: 5.84 min Scan# 1264

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10


Tgt Ion: 63 Resp: 1005137

Ion Ratio Lower Upper

63 100

98 6.7 3.3 9.9



100 4.3 2.1 6.2




Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 97.90 (97.60 to 98.60): VN0

Ion 99.90 (99.60 to 100.60): VN0









#25
 2-Butanone
 Concen: 265.172 ug/l
 RT: 6.82 min Scan# 1570
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampled :
 VSTDCCC050


Tgt Ion: 43 Resp: 1570984
 Ion Ratio Lower Upper
 43 100
 72 28.3 23.1 34.7



#26
 2,2-Dichloropropane
 Concen: 52.375 ug/l
 RT: 6.82 min Scan# 1568
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 77 Resp: 846587
 Ion Ratio Lower Upper
 77 100
 97 24.5 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 51.200 ug/l

RT: 6.82 min Scan# 1569

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050



Tgt Ion: 96 Resp: 659875

Ion Ratio Lower Upper

96 100

61 136.0 0.0 271.0



98 65.2 0.0 129.2



Abundance Ion 95.90 (95.60 to 96.60): VN0

Ion 60.90 (60.60 to 61.60): VN0

Ion 97.90 (97.60 to 98.60): VN0



#28

Bromochloromethane

Concen: 49.798 ug/l

RT: 7.18 min Scan# 1683

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10



Tgt Ion: 49 Resp: 432937

Ion Ratio Lower Upper

49 100

129 2.6 0.0 3.0


130 93.9 75.6 113.4




Abundance Ion 48.90 (48.60 to 49.60): VN0

Ion 128.80 (128.50 to 129.50): V

Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 258.150 ug/l
RT: 7.20 min Scan# 1687
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 42 Resp: 977686
Ion Ratio Lower Upper
42 100
72 47.0 37.6 56.4
71 44.4 35.6 53.4




Abundance Ion 42.00 (41.70 to 42.70): VN0
Ion 72.00 (71.70 to 72.70): VN0
Ion 71.00 (70.70 to 71.70): VN0






#30
Chloroform
Concen: 49.613 ug/l
RT: 7.36 min Scan# 1738
Delta R.T. -0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion: 83 Resp: 1021052
Ion Ratio Lower Upper
83 100
85 65.8 51.3 76.9



Abundance Ion 82.90 (82.60 to 83.60): VN0
Ion 84.90 (84.60 to 85.60): VN0





#31

Cyclohexane

Concen: 50.087 ug/l

RT: 7.65 min Scan# 1827

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion: 56 Resp: 862445

Ion Ratio Lower Upper

56 100

69 34.4 27.8 41.8




84 90.5 72.0 108.0



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 84.00 (83.70 to 84.70): VN0



#32

1,1,1-Trichloroethane

Concen: 51.795 ug/l

RT: 7.56 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10


Tgt Ion: 97 Resp: 901654

Ion Ratio Lower Upper

97 100

99 64.3 51.7 77.5



61 39.2 32.1 48.1




Abundance Ion 96.90 (96.60 to 97.60): VN0

Ion 98.90 (98.60 to 99.60): VN0

Ion 60.90 (60.60 to 61.60): VN0









#33
 1,2-Dichloroethane-d4
 Concen: 50.221 ug/l
 RT: 8.02 min Scan# 1942
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 65 Resp: 610400
 Ion Ratio Lower Upper
 65 100
 67 53.7 0.0 107.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.58 min Scan# 2117
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 114 Resp: 1739067
 Ion Ratio Lower Upper
 114 100
 63 18.2 0.0 37.0
 88 15.6 0.0 31.0









#35
Dibromofluoromethane
Concen: 50.150 ug/l
RT: 7.58 min Scan# 1806
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 113 Resp: 556875
Ion Ratio Lower Upper
113 100
111 103.0 81.4 122.2
192 24.2 18.8 28.2



#36
1,1-Dichloropropene
Concen: 52.133 ug/l
RT: 7.79 min Scan# 1870
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion: 75 Resp: 792650
Ion Ratio Lower Upper
75 100
110 38.0 19.1 57.1
77 30.8 25.2 37.8









#37
 Ethyl Acetate
 Concen: 51.077 ug/l
 RT: 6.91 min Scan# 1599
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 624196
 Ion Ratio Lower Upper
 43 100
 61 15.2 12.7 19.1
 70 11.7 9.0 13.4



#38
 Carbon Tetrachloride
 Concen: 50.185 ug/l
 RT: 7.77 min Scan# 1864
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 117 Resp: 810557
 Ion Ratio Lower Upper
 117 100
 119 97.4 77.8 116.6
 121 30.5 24.9 37.3









#39
Methylcyclohexane
Concen: 51.977 ug/l
RT: 9.08 min Scan# 2271
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 935675
Ion Ratio Lower Upper
83 100
55 73.4 57.0 85.4
98 46.5 37.2 55.8



#40
Benzene
Concen: 50.721 ug/l
RT: 8.03 min Scan# 1947
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion: 78 Resp: 2342845
Ion Ratio Lower Upper
78 100
77 23.8 18.2 27.4









#41
Methacrylonitrile
Concen: 53.062 ug/l
RT: 7.16 min Scan# 1675
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	266859
Ion	Ratio	Lower	Upper
41	100		
39	68.8	0.0	0.0#
67	105.7	0.0	0.0#
52	40.1	0.0	0.0#



#42
1,2-Dichloroethane
Concen: 50.315 ug/l
RT: 8.12 min Scan# 1973
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion:	62	Resp:	744750
Ion	Ratio	Lower	Upper
62	100		
98	10.4	0.0	20.0





#43

Isopropyl Acetate

Concen: 52.717 ug/l

RT: 8.15 min Scan# 1984

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :


MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 1060991




Ion	Ratio	Lower	Upper
43	100		
61	29.6	23.5	35.3
87	15.0	12.1	18.1



Abundance Ion 43.00 (42.70 to 43.70): VN057244.D (-1962) (-)

Ion 61.00 (60.70 to 61.70): VN057244.D (-1962) (-)

Ion 87.00 (86.70 to 87.70): VN057244.D (-1962) (-)



#44

Trichloroethene

Concen: 51.117 ug/l

RT: 8.83 min Scan# 2195


Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10



Tgt Ion: 130 Resp: 684429


Ion	Ratio	Lower	Upper
130	100		
95	94.1	0.0	181.8



Abundance Ion 129.80 (129.50 to 130.50): VN057244.D (-2177) (-)

Ion 94.90 (94.60 to 95.60): VN057244.D (-2177) (-)









#45
 1,2-Dichloropropane
 Concen: 52.097 ug/l
 RT: 9.11 min Scan# 2283
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 63 Resp: 609636
 Ion Ratio Lower Upper
 63 100
 65 30.4 25.0 37.4



#46
 Dibromomethane
 Concen: 49.792 ug/l
 RT: 9.20 min Scan# 2311
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 93 Resp: 392857
 Ion Ratio Lower Upper
 93 100
 95 85.4 67.0 100.4
 174 122.8 96.2 144.2





#47

Bromodichloromethane

Concen: 52.352 ug/l

RT: 9.40 min Scan# 2371

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion: 83 Resp: 796352

Ion Ratio Lower Upper




83 100

85 64.2 51.4 77.0

127 9.6 7.7 11.5



Abundance Ion 82.90 (82.60 to 83.60): VNC
Ion 84.90 (84.60 to 85.60): VNC
Ion 126.80 (126.50 to 127.50): VNC



#48

Methyl methacrylate

Concen: 54.084 ug/l

RT: 9.19 min Scan# 2308

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10


Tgt Ion: 41 Resp: 503660

Ion Ratio Lower Upper



41 100


69 94.0 75.7 113.5

39 52.8 43.3 64.9



Abundance Ion 41.00 (40.70 to 41.70): VNC
Ion 69.00 (68.70 to 69.70): VNC
Ion 39.00 (38.70 to 39.70): VNC









#49
1,4-Dioxane
Concen: 974.282 ug/l
RT: 9.19 min Scan# 2308
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 88 Resp: 204469
Ion Ratio Lower Upper
88 100
43 28.4 20.7 31.1
58 66.9 52.0 78.0



#50
Toluene-d8
Concen: 52.117 ug/l
RT: 10.09 min Scan# 2585
Delta R.T. -0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion: 98 Resp: 2110835
Ion Ratio Lower Upper
98 100
100 64.7 51.8 77.6





#51

4-Methyl-2-Pentanone

Concen: 255.792 ug/l

RT: 9.98 min Scan# 2552

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tgt Ion: 43 Resp: 3023004

Ion Ratio Lower Upper

43 100

58 42.4 33.7 50.5



Abundance Ion 43.00 (42.70 to 43.70): VN057244.D (-2535) (-)

Ion 58.00 (57.70 to 58.70): VN057244.D (-2535) (-)

2000000

1500000



1000000

500000

0

9.90 9.98 10.00

Time-->



#52

Toluene

Concen: 51.720 ug/l

RT: 10.15 min Scan# 2605

Delta R.T. -0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 92 Resp: 1484164

Ion Ratio Lower Upper

92 100

91 174.1 138.2 207.2



Abundance Ion 92.00 (91.70 to 92.70): VN057244.D (-2590) (-)

Ion 91.00 (90.70 to 91.70): VN057244.D (-2590) (-)

1500000


1000000


500000

0

10.10 10.15 10.20

Time-->





#53

t-1,3-Dichloropropene

Concen: 53.282 ug/l

RT: 10.38 min Scan# 2676

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tgt Ion: 75 Resp: 866942

Ion Ratio Lower Upper




75 100

77 31.3 25.3 37.9



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0



#54

cis-1,3-Dichloropropene

Concen: 53.553 ug/l

RT: 9.83 min Scan# 2507

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10


Tgt Ion: 75 Resp: 982041

Ion Ratio Lower Upper

75 100

77 32.3 25.0 37.6



39 40.3 32.3 48.5




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0









#55
 1,1,2-Trichloroethane
 Concen: 50.608 ug/l
 RT: 10.56 min Scan# 2732
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:	97	Resp:	567646
Ion	Ratio	Lower	Upper
97	100		
83	84.7	68.4	102.6
85	54.3	43.0	64.6
99	63.5	50.6	75.8



#56
 Ethyl methacrylate
 Concen: 52.764 ug/l
 RT: 10.43 min Scan# 2691
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion:	69	Resp:	830335
Ion	Ratio	Lower	Upper
69	100		
41	59.9	47.6	71.4
39	28.4	23.4	35.0





#57

1,3-Dichloropropane

Concen: 51.096 ug/l

RT: 10.70 min Scan# 2777

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 945576

Ion Ratio Lower Upper

76 100

78 32.3 26.0 39.0



#58

2-Chloroethyl Vinyl ether

Concen: 256.695 ug/l

RT: 9.69 min Scan# 2462

Delta R.T. 0.00 min

Lab File: VN057249.D




Acq: 13 Aug 2019 16:10


Tgt Ion: 63 Resp: 1183610

Ion Ratio Lower Upper

63 100

106 29.4 23.4 35.2









#59
 2-Hexanone
 Concen: 262.082 ug/l
 RT: 10.75 min Scan# 2791
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 2197622
 Ion Ratio Lower Upper
 43 100
 58 58.6 29.1 87.4



#60
 Dibromochloromethane
 Concen: 52.115 ug/l
 RT: 10.90 min Scan# 2838
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 129 Resp: 655523
 Ion Ratio Lower Upper
 129 100
 127 77.2 38.9 116.6









#61
 1,2-Dibromoethane
 Concen: 51.784 ug/l
 RT: 11.00 min Scan# 2870
 Delta R.T. -0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 107 Resp: 603009
 Ion Ratio Lower Upper
 107 100
 109 95.0 75.7 113.5



#62
 4-Bromofluorobenzene
 Concen: 51.740 ug/l
 RT: 12.40 min Scan# 3305
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion: 95 Resp: 718649
 Ion Ratio Lower Upper
 95 100
 174 94.2 0.0 187.6
 176 92.2 0.0 185.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2995

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion:117 Resp: 1553742

Ion Ratio Lower Upper

117 100

82 51.9 40.9 61.3




119 31.9 25.4 38.2



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#64

Tetrachloroethene

Concen: 49.140 ug/l

RT: 10.63 min Scan# 2754

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Tgt Ion:164 Resp: 674392


Ion Ratio Lower Upper

164 100

166 129.2 104.9 157.3

129 89.2 71.9 107.9

131 85.0 69.1 103.7






Abundance Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 51.101 ug/l
RT: 11.43 min Scan# 3003
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 1624698
Ion Ratio Lower Upper
112 100
114 32.3 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 50.780 ug/l
RT: 11.51 min Scan# 3027
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion:131 Resp: 623468
Ion Ratio Lower Upper
131 100
133 95.0 46.9 140.7
119 65.7 32.3 96.8









#67
Ethyl Benzene
Concen: 52.433 ug/l
RT: 11.51 min Scan# 3028
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 2845953
Ion Ratio Lower Upper
91 100
106 32.1 25.4 38.0



#68
m/p-Xylenes
Concen: 106.632 ug/l
RT: 11.62 min Scan# 3062
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion: 106 Resp: 2159443
Ion Ratio Lower Upper
106 100
91 201.2 161.4 242.0





#69

o-Xylene

Concen: 52.263 ug/l

RT: 11.95 min Scan# 3164

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampled :





VSTDCCC050

Tgt Ion:106 Resp: 1036298

Ion Ratio Lower Upper

106 100

91 213.6 106.5 319.4



#70

Styrene

Concen: 56.642 ug/l

RT: 11.96 min Scan# 3169

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10




Tgt Ion:104 Resp: 1738870


Ion Ratio Lower Upper

104 100

78 47.6 38.6 57.8

103 55.1 43.9 65.9









#71
Bromoform
Concen: 53.305 ug/l
RT: 12.13 min Scan# 3220
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:173 Resp: 501492
Ion Ratio Lower Upper
173 100
175 48.7 24.4 73.2
254 0.1 0.2 0.2#




Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion:152 Resp: 667608
Ion Ratio Lower Upper
152 100
115 55.1 27.6 82.8
150 174.8 0.0 348.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 54.929 ug/l

RT: 12.25 min Scan# 3258

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tgt Ion:105 Resp: 2817171

Ion Ratio Lower Upper




105 100

120 26.8 13.8 41.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 53.661 ug/l

RT: 12.07 min Scan# 3202

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Tgt Ion: 43 Resp: 840863


Ion Ratio Lower Upper

43 100

70 47.1 37.4 56.2

55 25.4 21.0 31.4

61 26.2 20.9 31.3






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V





#75

1,1,2,2-Tetrachloroethane

Concen: 54.541 ug/l

RT: 12.50 min Scan# 3337

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion: 83 Resp: 759636

Ion Ratio Lower Upper

83 100

131 11.2 5.6 16.8




85 63.8 32.3 96.8



Abundance Ion 82.90 (82.60 to 83.60): VN057249.D (-3243) (-)

Ion 130.80 (130.50 to 131.50): VN057249.D (-3243) (-)

Ion 84.90 (84.60 to 85.60): VN057249.D (-3243) (-)



#76

1,2,3-Trichloropropane

Concen: 84.285 ug/l

RT: 12.55 min Scan# 3353

Delta R.T. 0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 75 Resp: 958166

Ion Ratio Lower Upper



75 100


77 0.0 0.0 0.0



Abundance Ion 74.90 (74.60 to 75.60): VN057249.D (-3260) (-)

Ion 77.00 (76.70 to 77.70): VN057249.D (-3260) (-)





#77

Bromobenzene

Concen: 53.679 ug/l

RT: 12.53 min Scan# 3344

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion:156 Resp: 743767

Ion Ratio Lower Upper

156 100

77 175.5 88.0 264.1




158 98.2 49.4 148.0



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VN

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 54.283 ug/l

RT: 12.59 min Scan# 3364

Delta R.T. 0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 91 Resp: 3039759

Ion Ratio Lower Upper



91 100


120 24.0 12.1 36.3



Abundance Ion 91.00 (90.70 to 91.70): VN

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 54.508 ug/l

RT: 12.67 min Scan# 3390

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 1836809

Ion Ratio Lower Upper

91 100

126 36.2 18.0 54.0



#80

1,3,5-Trimethylbenzene

Concen: 54.567 ug/l

RT: 12.73 min Scan# 3409

Delta R.T. 0.00 min

Lab File: VN057249.D




Acq: 13 Aug 2019 16:10


Tgt Ion: 105 Resp: 2375553

Ion Ratio Lower Upper

105 100

120 49.8 25.3 75.8





#81

trans-1,4-Dichloro-2-butene

Concen: 48.817 ug/l

RT: 12.30 min Scan# 3273

Delta R.T. -0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion: 75 Resp: 242637

Ion Ratio Lower Upper

75 100

53 84.5 65.8 98.6




89 46.4 36.7 55.1



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#82

4-Chlorotoluene

Concen: 49.068 ug/l

RT: 12.77 min Scan# 3421

Delta R.T. 0.00 min

Lab File: VN057249.D


Acq: 13 Aug 2019 16:10

Tgt Ion: 91 Resp: 1786577

Ion Ratio Lower Upper



91 100


126 35.5 17.9 53.7



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 125.90 (125.60 to 126.60): V









#83
 tert-Butylbenzene
 Concen: 54.786 ug/l
 RT: 12.99 min Scan# 3490
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:119 Resp: 2108015
 Ion Ratio Lower Upper
 119 100
 91 62.1 30.8 92.3
 134 25.6 12.9 38.7



#84
 1,2,4-Trimethylbenzene
 Concen: 48.174 ug/l
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion:105 Resp: 2264780
 Ion Ratio Lower Upper
 105 100
 120 45.8 23.2 69.5





#85

sec-Butylbenzene

Concen: 54.091 ug/l

RT: 13.17 min Scan# 3546

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:105 Resp: 2655698

Ion Ratio Lower Upper

105 100

134 20.9 10.4 31.2



#86

p-Isopropyltoluene

Concen: 48.032 ug/l

RT: 13.29 min Scan# 3582

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10




Tgt Ion:119 Resp: 2358001


Ion Ratio Lower Upper

119 100

134 27.2 13.5 40.5

91 22.1 11.1 33.3









#87
 1,3-Dichlorobenzene
 Concen: 51.322 ug/l
 RT: 13.28 min Scan# 3580
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:146 Resp: 1186971
 Ion Ratio Lower Upper
 146 100
 111 37.2 18.8 56.4
 148 64.4 32.1 96.5



#88
 1,4-Dichlorobenzene
 Concen: 49.830 ug/l
 RT: 13.36 min Scan# 3605
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion:146 Resp: 1120929
 Ion Ratio Lower Upper
 146 100
 111 37.3 18.3 54.8
 148 65.1 32.0 96.0









#89
n-Butylbenzene
Concen: 50.478 ug/l
RT: 13.62 min Scan# 3683
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10




Instrument :
MSVOA_N
ClientSampled :
VSTDCCC050


Tgt Ion: 91 Resp: 1738175
Ion Ratio Lower Upper
91 100
92 51.7 26.3 78.9
134 28.0 14.1 42.3



#90
Hexachloroethane
Concen: 52.350 ug/l
RT: 13.88 min Scan# 3764
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion: 117 Resp: 427648
Ion Ratio Lower Upper
117 100
201 98.9 49.5 148.5





#91

1,2-Dichlorobenzene

Concen: 51.881 ug/l

RT: 13.65 min Scan# 3695

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tgt Ion:146 Resp: 1169889

Ion Ratio Lower Upper

146 100

111 38.6 19.3 57.9




148 63.1 31.9 95.5



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 53.872 ug/l

RT: 14.27 min Scan# 3887

Delta R.T. 0.00 min

Lab File: VN057249.D

Acq: 13 Aug 2019 16:10


Tgt Ion: 75 Resp: 120138

Ion Ratio Lower Upper

75 100

155 100.2 50.3 151.0



157 130.3 65.3 195.9




Abundance Ion 74.90 (74.60 to 75.60): V

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 58.412 ug/l
 RT: 14.91 min Scan# 4086
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tgt Ion:180 Resp: 518447
 Ion Ratio Lower Upper
 180 100
 182 95.5 47.7 143.1
 145 28.8 14.2 42.6




Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.80 (144.50 to 145.50): V






#94
 Hexachlorobutadiene
 Concen: 57.525 ug/l
 RT: 15.01 min Scan# 4117
 Delta R.T. 0.00 min
 Lab File: VN057249.D
 Acq: 13 Aug 2019 16:10

Tgt Ion:225 Resp: 489340
 Ion Ratio Lower Upper
 225 100
 223 61.6 31.4 94.2
 227 64.9 31.4 94.3



Abundance Ion 224.70 (224.40 to 225.40): V
 Ion 222.70 (222.40 to 223.40): V
 Ion 226.70 (226.40 to 227.40): V









#95
Naphthalene
Concen: 60.483 ug/l
RT: 15.13 min Scan# 4153
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 892186
Ion Ratio Lower Upper
128 100
127 12.6 10.4 15.6
129 10.8 8.5 12.7



#96
1,2,3-Trichlorobenzene
Concen: 61.401 ug/l
RT: 15.30 min Scan# 4206
Delta R.T. 0.00 min
Lab File: VN057249.D
Acq: 13 Aug 2019 16:10

Tgt Ion:180 Resp: 493037
Ion Ratio Lower Upper
180 100
182 94.6 48.9 146.7
145 29.9 15.0 45.1

