

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081319\
 Data File : VN057256.D
 Acq On : 13 Aug 2019 19:24
 Operator : JC/SP
 Sample : K4290-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-1

Quant Time: Aug 14 12:21:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 06:25:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1429581	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1977981	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1577791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	392551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	657706	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
35) Dibromofluoromethane	7.58	113	593039	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
50) Toluene-d8	10.09	98	2243190	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
62) 4-Bromofluorobenzene	12.40	95	585774	37.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.16%	
Target Compounds						
16) Acetone	3.82	43	16614	3.360	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081319\
Data File : VN057256.D
Acq On : 13 Aug 2019 19:24
Operator : JC/SP
Sample : K4290-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-1

Quant Time: Aug 14 12:21:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 06:25:03 2019
Response via : Initial Calibration

