

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081319\
 Data File : VN057257.D
 Acq On : 13 Aug 2019 19:48
 Operator : JC/SP
 Sample : K4290-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-4

Quant Time: Aug 14 07:19:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 06:25:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1480157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2062528	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1694572	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	455495	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	677852	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
35) Dibromofluoromethane	7.58	113	624240	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	10.09	98	2353278	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
62) 4-Bromofluorobenzene	12.40	95	642225	38.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.98%	
Target Compounds						
16) Acetone	3.82	43	16176	3.160	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081319\
Data File : VN057257.D
Acq On : 13 Aug 2019 19:48
Operator : JC/SP
Sample : K4290-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-4

Quant Time: Aug 14 07:19:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 06:25:03 2019
Response via : Initial Calibration

