

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081319\
 Data File : VN057262.D
 Acq On : 13 Aug 2019 21:50
 Operator : JC/SP
 Sample : K4266-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R1-R2-A2-(0)

Quant Time: Aug 14 07:37:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 06:25:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1482393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2050366	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1700644	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	421174	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	687045	45.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.88%	
35) Dibromofluoromethane	7.58	113	620291	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	10.09	98	2377510	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
62) 4-Bromofluorobenzene	12.40	95	621120	37.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.86%	
Target Compounds						
16) Acetone	3.81	43	31922	6.226	ug/l #	85
20) Methylene Chloride	4.55	84	47717	3.186	ug/l	96
43) Isopropyl Acetate	8.16	43	190265	8.018	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081319\
Data File : VN057262.D
Acq On : 13 Aug 2019 21:50
Operator : JC/SP
Sample : K4266-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R1-R2-A2-(0)

Quant Time: Aug 14 07:37:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 06:25:03 2019
Response via : Initial Calibration

