

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081319\
 Data File : VN057269.D
 Acq On : 14 Aug 2019 00:41
 Operator : JC/SP
 Sample : K4266-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R1-R2-C2-(4.3)

Quant Time: Aug 14 07:49:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 06:25:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1465580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2036228	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1707376	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	440738	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	675667	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
35) Dibromofluoromethane	7.58	113	611272	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
50) Toluene-d8	10.09	98	2379397	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
62) 4-Bromofluorobenzene	12.40	95	624851	38.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.84%	
Target Compounds						
16) Acetone	3.81	43	32039	6.321	ug/l	93
18) Methyl Acetate	4.31	43	54469	3.076	ug/l #	69
20) Methylene Chloride	4.55	84	48779	3.294	ug/l	95
43) Isopropyl Acetate	8.16	43	315794	13.401	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081319\
Data File : VN057269.D
Acq On : 14 Aug 2019 00:41
Operator : JC/SP
Sample : K4266-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R1-R2-C2-(4.3)

Quant Time: Aug 14 07:49:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 06:25:03 2019
Response via : Initial Calibration

