

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081320\
 Data File : VN062906.D
 Acq On : 13 Aug 2020 13:13
 Operator : JC/MD
 Sample : L3679-14
 Misc : 5.00µ/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-081220

Quant Time: Aug 14 05:42:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	130411	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	226748	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	225288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	93684	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.98	65	106725	49.49	ug/l	0.00
Spiked Amount			Recovery	=	98.98%	
35) Dibromofluoromethane	7.54	113	78853	50.84	ug/l	0.00
Spiked Amount			Recovery	=	101.68%	
50) Toluene-d8	10.06	98	295467	48.79	ug/l	0.00
Spiked Amount			Recovery	=	97.58%	
62) 4-Bromofluorobenzene	12.38	95	108469	47.40	ug/l	0.00
Spiked Amount			Recovery	=	94.80%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062906.D
Acq On : 13 Aug 2020 13:13
Operator : JC/MD
Sample : L3679-14
Misc : 5.00µ/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-081220

Quant Time: Aug 14 05:42:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

