

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081320\
 Data File : VN062908.D
 Acq On : 13 Aug 2020 14:02
 Operator : JC/MD
 Sample : L3619-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-6

Quant Time: Aug 14 05:46:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	126295	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	220230	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	232425	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	83408	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	109198	52.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.56%	
35) Dibromofluoromethane	7.55	113	75657	50.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.44%	
50) Toluene-d8	10.06	98	283750	48.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.48%	
62) 4-Bromofluorobenzene	12.38	95	112437	50.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062908.D
Acq On : 13 Aug 2020 14:02
Operator : JC/MD
Sample : L3619-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-6

Quant Time: Aug 14 05:46:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

